

An ultra-high performance liquid chromatography (UHPLC) based method for the analysis of d-xylonate was developed using an amide column in combination with an evaporative light scattering (ELS) detector. Separation of d-xylonate from other components of the fermentation medium was achieved. The dynamic range of the method was 0.2-7.0 g/L.

